




检验报告



CTTT-WT20010048

客户提供信息 Information Provided by Client	委托单位/地址 Applicant	江苏友奥医疗科技有限公司 YOAU Medical Technology Co., Ltd 江苏省常州市武进区礼嘉镇甘棠路10号 No 10, Gantang Road, Lijia Town, Wujin District, Changzhou City, Jiangsu Province			送样人: Contact	葛莉莉 Ge Lili
	生产单位/地址 Manufacturer	江苏友奥医疗科技有限公司 YOAU Medical Technology Co., Ltd 江苏省常州市武进区礼嘉镇甘棠路10号 No 10, Gantang Road, Lijia Town, Wujin District, Changzhou City, Jiangsu Province				
	样品信息 Information of Submitted Sample	样品名称: Sample Name	KN95防护口罩 KN95 Protective Mask	商标: Trademark	YOAU	
		样品总数: Sample Count	20只 20Piece (s)	颜色: Colour	-----	
		号型规格: Size	-----	安全类别: Safety Category	-----	
		质量等级: Quality Grade	-----			
		产品款号或货号: Style No. or Order No.	-----			
样品描述 Test Part Description	1# 白色口罩 White Mask					
检验性质 Test Type	委托检验 Commission Test	样品受理日期 Date of Submission	2020-05-09	报告签发日期 Date of Checking	2020-05-12	
执行标准 Test Standards	GB 2626-2006 呼吸防护用品 自吸过滤式防颗粒物呼吸器 Respiratory protective equipment—Non-powered air-purifying particle respirator					
检验结论 Conclusion	检验结果及符合性见附页。 Test results and compliance refer to next page(s). 检验单位盖章Stamp of Inspection Unit					
非标准检验方法说明 Non-standard Check Methods	-----					
检验结果的不确定度 Uncertainty of Result	-----					
备注 Remarks	客户要求呼吸阻力、可燃性、过滤效率、外观检查不做预处理, 过滤效率测试初始过滤效率, 所检项目按GB 2626-2006判定。 As per client's request, Respiratory Resistance, Flammability, Filtration Efficiency and Appearance Inspection are not pretreated. The Filtration Efficiency is required to test only initial filtration efficiency. The tested items are judged according to the standard GB 2626-2006. CNAS (L0608) 授权过滤效率项目, 未授权呼吸阻力、可燃性、外观检查和视野项目。 CMA (200215340010) 授权所有所检项目。 CNAS (L0608) authorizes the test item Filtration Efficiency, the test items Respiratory Resistance, Flammability, Appearance Inspection and View are unauthorized by CNAS (L0608). CMA (200215340010) authorizes all the test items.					



批准:
Approver

方倩

审核:
Checker

张涛

编制:
Editor

岳玉婷



检验报告



中国认可
国际互认
检测
TESTING
CNAS L0008

CTTT-WT20010048

检测项目 Test Items	项目描述 Description	单位 Unit	标准值 Standard Requirement	实测值 Results	评价 Conclusions	方法标准/备注 Test Method/ Remarks
1#						
呼吸阻力 Respiratory Resistance	总呼气阻力 Total Expiratory Resistance	Pa	≤250	未预处理: Without Pretreatment: 98	符合 Pass	GB 2626-2006
	总吸气阻力 Total Inspiratory Resistance	Pa	≤350	未预处理: Without Pretreatment: 130		
可燃性 Flammability	---	---	暴露于火焰的各部件在从火焰移开后, 不应燃烧; 如果燃烧, 续燃时间不应超过5s Each part exposed to the flame shall not burn after it is removed from the flame; if it burns, the afterflame time shall not exceed 5s.	未预处理: Without Pretreatment: 未继续燃烧 No continue burning	符合 Pass	GB 2626-2006
过滤效率 Filtration Efficiency (KN95)	氯化钠颗粒物 NaCl Particle	%	≥95.0	未预处理: Without Pretreatment: 初始过滤效率: Initial Filtration Efficiency: 98.6	符合 Pass	GB 2626-2006
外观检查 Appearance Inspection	样品表面 Surface of The Sample	---	按标准要求 As per standard requirement	+	符合 Pass	GB 2626-2006
	头带 Head Harness	---	按标准要求 As per standard requirement	+		



检验报告



CTTT-WT20010048

检测项目 Test Items	项目描述 Description	单位 Unit	标准值 Standard Requirement	实测值 Results	评价 Conclusions	方法标准/备注 Test Method/ Remarks
视野 View	下方视野 View Below Mask	°	≥60	61	符合 Pass	GB/T 2891-1995

表中“+”表示符合标准要求，“X”表示不符合标准要求。

+ Meet the standard requirements, X Not Meet the standard requirements.

以下空白
Blank Part Below

TTTS

TTTS

TTTS

TTTS

TTTS

TTTS

TTTS

TTTS

TTTS



天纺标检测认证股份有限公司



200215340010

检验报告



中国认可
国际互认
检测
TESTING
CNAS L0608

第4页 共4页

Page 4 of 4

CTTT-WT20010048

样 品 Sample

